

Spectrophotometer with excellent repeatability accuracy




Ultra-high repeatability accuracy: dE*ab ≤ 0.02



Spectrophotometer DigiSpec®

DS-600 | DS-620 | DS-660

Key technology



Dual optical path design improves repeatability accuracy $dE^*ab \le 0.02$

The dual optical path design monitors light source energy fluctuations while measuring the sample signal, reducing interference during measurement, obtaining higher measurement stability and improving the instrument's measurement repeatability index to dE*ab \leq 0.02. This ensures the instrument, high standards of measurement speed, accuracy, stability and inter-stage difference consistency. The relevant technology is protected by a Chinese invention patent and a US invention patent.

Innovative 5 micron thick nano-integrated optics

Innovation is the soul of CHNSpec. After nearly 10 years of dedicated research, CHNSpec has adopted nano-integrated optics as the spectroscopic device, which can achieve nanolevel spectroscopic capability with only 5 micron thickness of optics, once again leading the direction of innovation in the industry, crossing the technological blockade of foreign



products and greatly enhancing the technical performance of the products. The relevant technology is protected by Chinese invention patents.

- Related techniques were published in the Chinese famous optical academic journals
 - "Journal of Optics" and "Journal of Photonics"
 - "Optimized Design of Spectrophotometer Based on Real-Time Dual Optical Correction"
 - "The Design of SCE Measurement Gloss Correction Model for Color Measurement Instruments Based on D/8 Condition"
- The technology is protected by a Chinese invention patent:

Colour measuring instrument and implementation method based on D/8 condition for

light trap error correction CN

SAME US9243953B1

CN201310373360.1

A colorimeter for measuring the colour of objects using linear variable filters

CN201310027285.3

- Related technologies are protected by US invention patents: SPECTROPHOTOMETRIC COLORIMETER BASED ON LED LIGHT SOURCE AND METHOD FOR REALIZING THE
- The relevant technology won the third prize of Science and Technology Progress of Zhejiang Province and the Excellent Product Award of China Instrument Society



Product features

Ultra-high repeatability accuracy: dE*ab ≤ 0.02

Repeatability accuracy is an important indicator to describe the accuracy of spectrophotometers. The DS-600 series spectrophotometers are evaluated on the basis of a rigorous standard of repeatability, which is unmatched by any other spectrophotometer in its class.



Excellent inter-stage consistency

The superior level of technology and craftsmanship ensures that the DS600 series has excellent inter-instrument agreement. The use of BCRA series standard colour bricks for colour transfer and colour value traceability guarantees an excellent level of inter-stage variation.

Over 30 measurement parameters and nearly 40 evaluation light sources available

The DS-600 series spectrophotometers offer 30+ measurement indicators such as spectral reflectance, CIE-Lab, CIE-LCh, ΔE^* ab, opacity, whiteness, yellowness, etc.; and nearly 40 evaluation light sources to choose from such as A, B, C, D50, D55, D65, etc., covering almost all colour measurement indicators and light source types in the industry.

Calibrated white plate (artificial diamond zirconium material)

- ·Mohs hardness: 9
- ·Spectral reflectance >90%
- •No discolouration due to changes in temperature and humidity
- ·No discolouration by oxidation
- ·Ultra-high strength without scratching





Calibration base and zirconium reference with a Mohs hardness of 9 to calibrate the instrument, ensuring long-term stability

Compared to existing products, the DS-600 series spectrophotometer does not require frequent manual calibration when in use. Simply place it on the calibration base and the instrument will automatically calibrate the overall instrument function and accuracy according to its own state and environmental factors, ensuring that the instrument is always in a stable state and ready for use.

The white plate in the calibration base is the basis of the instrument's work. Through long-term investment and research, CHNSpec has integrated the "artificial diamond"

zirconium material as the calibration white plate, with a Mohs hardness of 9. As the material itself has the hardness and stability comparable to diamond, the surface of the calibration white plate will not be scratched and will not change colour with changes in temperature and humidity. This is a further improvement in the stability and durability of the calibration whiteboard compared to similar foreign and domestic products that use common industrial ceramics or even plastic as calibration whiteboards, ensuring the performance of the instrument.



Spectrophotometer DigiSpec®

DS-600 | DS-620 | DS-660

Product features

The DS-600 series supports 6 measuring apertures for selection

In order to facilitate users to measure samples of different sizes, the DS-600 series spectrophotometer supports 6 apertures for customers to use, stable type: Φ 11mm+ Φ 6mm+ Φ 3mm, dexterous type: ∇ 11mm+ ∇ 6mm+ ∇ 3mm, flexible for various use and testing conditions.









Built-in HD camera for clear observation of the measured area

The DS-600 series spectrophotometer can obtain an image of the measured area through the camera when measuring, which can clearly locate the measured area of the sample and avoid inaccurate measurement due to wrong area.

Support WeChat applet, Android, Apple, Hongmeng mobile APP

- The DS-600 series spectrophotometers can be connected to a variety of mobile phones via a rich mobile app.
- Users no longer have to pass on the colour values of samples and physical objects, they can easily pass on colour data via WeChat.
- Users can find the most similar colours in multiple sets of colour cards.
- Users can create personal colour databases and enter information on colour cards for printing, paint and textiles. The
 colour libraries created can be uploaded to the cloud for easy colour processing with data sharing across multiple
 devices.
- Business users can create and manage their own colour card information library and colour recipes in the cloud, and share the information library and colour recipes to their own users through a unique invitation code.



Check color card number



Set standard with color



Share color data



WechatApp



APP Mobile APP

Spectrophotometer DigiSpec®

DS-600 | DS-620 | DS-660

Product features

Use the powerful PC-based colour management system ColorExpert*








The DS-600 series spectrophotometer is supplied with a Windows colour management system, ColorExpert, which connects to the DS-600 series spectrophotometer via Bluetooth or USB cable.

ColorExpert is a full-featured colour management software with four functional modules: My Colours, Colour Check, Colour Matching System and Personal Centre.



My Colour

Users can collect or create their own colour library from hundreds of other users' own shared colour libraries. The PC software and mobile app can share accounts and the colour library data follows the account to synchronise information between PC and mobile.




Colour QC

The user can calibrate, measure and set up the spectrophotometer via the PC software. The user can use colours from the cloud database as specimens to measure colour differences, view spectrograms, colour difference graphs, specimen trial data and export the desired data test reports.



Color matching

Provide users with a more convenient and efficient color matching process. After the instrument measures the color of the sample, the system calculates the formula in the formula center and automatically fixes the color to achieve accurate matching. The system is suitable for computer automatic color matching applications in paint, coating, printing, textile and other industries.



Personal centre

Users can edit personal information, search for or delete connected instruments, manage downstream users, and manage color libraries shared with downstream users.



Function difference



Model	DS-600	DS-620	DS-660
Test condition	SCI+SCE	SCI+SCE	SCI+SCE
Repeatability	≤ 0.04	≤ 0.03	≤ 0.02
Inter-instrument Agreement	≤ 0.4	≤ 0.3	≤ 0.2
Aperture	2 apertures	6 apertures	6 apertures
UV light source	×	✓	✓
Camera function	×	✓	✓
Mobile App	×	✓	✓
PC software	✓	✓	✓



Technical Pa	rameter				
Product model	DS-600	DS-620	DS-660		
Measuring structure*		D/8, SCI/SCE			
Measurement repeatability**	ΔE*ab≤ 0.04	ΔE*ab≤ 0.03	ΔE*ab≤ 0.02		
Inter-instrument Agreem	nent*** ΔE*ab≤0.4	ΔE*ab≤0.3	ΔE*ab≤0.2		
Display accuracy		0.01			
Measuring aperture	2 apertures: Stable type: Φ11mm Dexterous type: ∇11mm	6 apertures: Stable type: Φ11mm,Φ6mm,Φ3mm Dexterous type: ▽11mm,▽6mm,▽3mm			
Color Spaces and Indices	Reflectance, CIE-Lab, CIE-LCh, HunterLab, CIE Luv, XYZ, Yxy, RGB, Color difference(ΔΕ*ab, ΔΕ*cmc, ΔΕ*94,ΔΕ*00),WI(ASTM E313-00,ASTM E313-73,CIE,ISO2470/R457, AATCC, Hunter, Taube Berger Stensby), YI(ASTM D1925,ASTM E313-00, ASTM E313-73), Blackness(My,dM),Color Fastness				
Source condition	A,B,C,D50,D55,D65,D75,F1,F2,F3,F4,F5,F6,F7,F8,F9,F10,F11,F12,CWF,U30,U35,DLF,NBF,TL83, TL84,ID50,ID65,LED-B1,LED-B2,LED-B3,LED-B4,LED-B5,LED-BH1,LED-RGB1,LED-V1,LED-V2				
Light source	LED	LED+UV			
Measurement observation method	Visual	Camera			
Calibration		Auto calibration			
Software support	Windows	Andriod,iOS,Windows, Wechat app			
Guaranteed accuracy	Guaranteed measurement	Guaranteed first class measurement			
Observer	2°, 10°				
Integrating sphere diame					
Standards	CIE No.15,GB/T 3978,	GB 2893,GB/T 18833,ISO7724-1,AST	M E1164,DIN5033 Teil7		
Ways of spectral	Nano-integrated spectral devices				
Sensor	Silicon photodiode array Dual 16-group				
Wavelength interval	10nm				
Wavelength range	400-700nm (user viewable reflectance at 31 wavelengths)				
Reflectance determination	•				
Reflectance resolution	0.01%				
Measurement method	Single measurement, average measurement (2 to 99 measurements)				
Measurement time		Approx. 1 second			
Interface	USB	USB, E	lluetooth		
Screen		Screen Full colour screen, 3.5			
Battery capacity	8000 continuous measurements on a single charge				
Life of light	10 years and 1 million cycles				
Language	Simplified Chinese, English				
Storage	10,000 data items	Instrument :10,000 data			

 $^{^{\}star}\, \text{Diffuse illumination} \, / \, 8^{\circ}\, \text{directional reception with specular reflected light included} \, / \, \text{specular reflected light removed}$

^{**}White plate calibration with 30 standard deviations measured at 5 second intervals after white plate calibration

^{***}BCRA Series II average of 12 colour plate measurements